

Laboratory			Al Bader		LABTECH		PROMPT	
No	Calibration field	Calibration item	Range	CMC	Range	CMC	Range	CMC
04	Thermal	Glass thermometer	-40 to 200°C 25 to 650 °C	0.7°C 1.6°C			-30°C to 660°C	0.25 °C
		Prop thermometer, switch gauge, transmitter	-40 to 200°C 25 to 650 °C	0.7°C 1.6°C	0 to 250°C /0.01 ° c	0.11° c	-30°C to 660°C	0.25 °C
		Controllers,	-40 to 200°C/0.01°C 25 to 650°C/0.01°C	0.7°C 1.6°C	-30 to 1600°C /0.01°C -200 to 850°C /0.01°C	0.02°C 0.1°C	-30°C to 660°C	0.25 °C
		Temperature blocks			-40 to 400°C /0.1°C Up to 1000°C /0.1°C	0.13°C 1.20°C		
		Thermocouple Indicators Type K	-40 to 200°C/0.01°C 25 to 650°C/0.01°C	0.7°C 1.6°C	-200 to 1700°C /0.01° C	0.1° C	-250°C to 1372°C	0.46 °C

Evrakın elektronik imzalı suretine <http://e-belge.gtb.gov.tr> adresinden 23898c5e-6541-4ee6-88e3-387b2481595c kodu ile eri ebilirsiniz.
 BELGEN N ASLI ELEKTRON K MZALIDIR.