

Laboratory			Al Bader		LABTECH		PROMPT	
No	Calibration field	Calibration item	Range	CMC	Range	CMC	Range	CMC
04	Thermal	Ovens freezers Fridge incubators	Up to 600°C/0.1°C To 40°C/0.1°C  Up to100°C/0.1°C	0.75°C 0.6°C  0.6°C	30 to 250°C  0 to 20°C /0.1°C 0 TO 90 °C	0.70 °C  0.70 °C 0.70°C	-195°C to 419°C	0.63°C
		Liquid baths	Up to 600°C/0.1°C	0.75°C	5 to 95 °C /0.1°C	0.32°C		
		Muffle furnaces	Up to 1200°C/1°C	4.45°C	300 to 1000°C	1.20°C		
		Humidity equipment	0% R.H to 95% R.H	3% R.H				
05	Volume	Pipette, micropipette	5 to 1000µL/0.1µL 1mL to 2000mL/0.1mL	3.1µL  5.66mL				

Evrakın elektronik imzalı suretine <http://e-belge.gtb.gov.tr> adresinden 23898c5e-6541-4ee6-88e3-387b2481595c kodu ile erişebilirsiniz.  
BELGEN N ASLI ELEKTRONİK MZALIDIR.