

Laboratory			Al Bader		LABTECH		PROMPT	
No	Calibration field	Calibration item	Range	CMC	Range	CMC	Range	CMC
06	Dimensional	Calipers	0 to 300mm/0.01 mm >300 to 600mm/0.01 mm	12µm 13µm	0 to 150mm /0.01mm	0.007 mm	0 mm to 300mm 0 mm to 300mm	9.0µm 17.8 µm
		Micrometers	0to150mm/0.01mm >150to300mm /0.01mm >300 to 600mm/0.01mm	1.4µm 6.8µm 7.8µm	0 to 25 mm /0.001 mm	0.007 mm	0 mm to 150 mm 0 mm to 150 mm	6.0 µm 1.9µm
		Height gages-length Dial Digital Vernier	0 to 300mm/0.01 mm >300 to 600mm/0.01 mm	8 µm 13µm	0 to 100 mm /0.001 mm	0.0002 mm	0 mm to 600mm 0 mm to 600mm	11.0µm 15.0µm
		Gage blocks-length						
		Cylindrical gage -plugs and pins	Up to 10mm Up to 150mm	3.2µm 3.5µm	0.5to 300mm	0.0012 mm		
		Thread plug gages Pitch diameter Major diameter						

Evrağın elektronik imzalı suretine <http://e-belge.gtb.gov.tr> adresinden 23898c5e-6541-4ee6-88e3-387b2481595c kodu ile eri ebilirsiniz.
BELGEN N ASLI ELEKTRON K MZALIDIR.