

Laboratory			Al Bader		LABTECH		PROMPT	
No	Calibration field	Calibration item	Range	CMC	Range	CMC	Range	CMC
06	Dimensional	Ring gauges-cylindrical &tapered Bore gauge	Up to 2mm	7µm	0 to 200 mm	0.0002m m		
		Length indicators						
		Length standard			0 to 5000 mm /0.5 mm	0.12 mm		
		Metal tape measure	0mm to 1000mm /0.5mm 0mm to 5000mm /1mm	0.6 µm 1.7 mm	0 to 50000 mm /0.5 mm	0.12 mm		
		Coating thickness	Up to1507.15µm Up to 1503.5 µm	1.4 µm 1 µm				
		Surface plates flatness						
		Thickness gauge Feeler gauge	Up to 50mm/0.01 mm Up to 2 mm	7µm 3.3µm	0 to 100 mm /0.001 mm 0.02 to 1mm	0.0002 mm 0.0011 mm	0 mm to 5 mm 0 mm to 25 mm 0 mm to 50 mm 0.01 mm to 2.0 mm	1.0 µm 5.9 µm 6.0 µm 2.7 µm

Evrakın elektronik imzalı suretine <http://e-belge.gtb.gov.tr> adresinden 23898c5e-6541-4ee6-88e3-387b2481595c kodu ile eri ebilirsiniz.
 BELGEN N ASLI ELEKTRON K MZALIDIR.