

Laboratory			Al Bader		LABTECH		PROMPT	
No	Calibration field	Calibration item	Range	CMC	Range	CMC	Range	CMC
07	Optical	Optical power measure (6 to 60)Db (10 to 110)dB (10 to 110)dB						
		Optical wavelength measure						
		Fiber Optical wavelength measure						
08	Time/frequency	Stopwatch/timer	Up to24h /0.01 sec	0.3 sec				
		Rotational speed/ Tachometer	1 to 11000 rpm/1rpm	2.5 rpm				
09	Radiation	Nuclear density gauge - density	1775.4 kg/m³ 2101.5kg/m³ 2143.1 kg/m³ 2174.3 kg/m³ 2179.1 kg/m³ 2188.8 kg/m³ 2226.8 kg/m³ 2685.7 kg/m³	0.3 % 1.3 % 1.1 % 1 % 1.2% 1% 0.3 % 0.3%				
		Nuclear density gauge - moisture	531.2 kg/m³ 546.5 kg/m³ 549.5 kg/m³ 564.3 kg/m³ 566.9 kg/m³ 592.5 kg/m³	4 % 4.5% 4% 1.3% 4.9% 4.1%				

Evrakın elektronik imzalı suretine <http://e-belge.gtb.gov.tr> adresinden 23898c5e-6541-4ee6-88e3-387b2481595c kodu ile erişebilirsiniz.

BELGENİN ASLI ELEKTRONİK İMZALIDIR.